


Air Sampling System MBASS30v3

Company	
Department	
Production Site	
Address	
ZIP Code / Town	

Manufacturer	Umweltanalytik Holbach GmbH, D-66687 Wadern
IQ/OQ-Protocol prepared	Signature: 
IQ/OQ-Protocol co-approved:	Signature:
This document is only valid if it is signed off by the manufacturer	

Extract only

Air Sampling System MBASS30v3



Air Sampling System MBASS30v3

Table of Contents

Table of Contents	3
1 IQ Purpose	4
1.1 IQ Responsibilities	4
1.2 Manufacturer	4
2 IQ MBASS30V3 with Air Sampling Head LKS100	4
2.1 Delivery Verification and Part Description	4
2.2 Documentation of Delivery Verification and Part Description	5
2.3 Activating and Conditioning of the Measuring Station	6
3 OQ Purpose	7
3.1 OQ Procedure	7
3.2 Verification of the Test Procedure	7
3.2.1 Sampling	8
3.2.2 After Sampling	9
3.3 Additional Information	10
3.3.1 Sampling Interruption	10
3.3.2 Sampling Disruptions	10
4 PQ Purpose	11
4.1 PQ – First Steps	11
4.1.1 Device Settings	11
4.1.2 Sampling Settings	13
4.1.3 Sampling presets	14

Extract only

Air Sampling System MBASS30v3

2.3 Activating and Conditioning of the Measuring Station

Before going operational, follow the steps listed below in the next table and sign it.

No.	Step	Corresponding section*	done/yes	Remarks	Date/Initials
1	Place MBASS30v3 safely at the location for sampling. Take care to position it securely and keep a minimum distance between the back and the wall.	7	<input type="checkbox"/>		
2	Turn the carrying handle into the front position .	4.2	<input type="checkbox"/>		
3	Carefully remove the protective foil from the display.		<input type="checkbox"/>		
4	Switch on MBASS30v3 on by pressing the On/Off button on the back side.	7.1	<input type="checkbox"/>		
5	The preset selection is shown on the screen.	7.1	<input type="checkbox"/>		
6	Select an appropriate preset program for the sampling.	7.2	<input type="checkbox"/>		
7	Optional: Edit one of the preset programs to fit your requirements (section 4.1.3).	8	<input type="checkbox"/>		

* For further details, please check the respective section of the operating manual

Verified by: Name of tester		Initials		Date	
Verified by: Name of supervisor		Initials		Date	

Air Sampling System MBASS30v3

3.3 Additional Information

3.3.1 Sampling Interruption

No.	Step	Corresponding section*	Remarks
1	The sampling can be interrupted by tapping the "Pause" button and be continued with the "Resume" button. It can also always be aborted by tapping the "Stop" button and confirmation by tapping the "Yes" button.	7.4, 7.6	

* For further details, please check the respective section of the operating manual

3.3.2 Sampling Disruptions

No.	Occurrence	Corresponding section*	Remarks
1	Low battery voltage	7.7	
2	The air intake is blocked, the volume flow cannot be maintained	7.7	
3	The flow sensor reports inexplicable values	7.7	
4	The fan motor does not start	7.7	
<p>In all these cases sampling is terminated and the alarm is signaled visually on the display and acoustically with a 2-tone sequence. The reason for the alarm is shown on the screen. Further information on the error can be shown by tapping the "i" button on the right.</p>			

* For further details, please check the respective section of the operating manual